

JAN 06 2010

Supervisor/Operator Performing the Verification Procedure: DataMaster cdm S/N 130289

Name Byron D. Redburn ID# 2087 Date 1-2-10

A Agency Nome Police Dept Phone # 443-5262

Instrument Location Nome Police Dept

B Alco S/N X301548 Target Value .079 High Pressure 400

Alco Test Values .079 .078

1st Alco 2nd Alco

Signature [Signature]

(OVER)

(Do Not write in the area below)

AMB
1/13/10


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 20th day of January, 2010.
[Signature] (Notary Seal Stamp)

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office
BT9 06/03



DataMaster cdm S/N 130289

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130289

JANUARY 02, 2010

OPERATOR'S NAME:
REDDURN BYRON D
OPERATOR'S NUMBER: 2257
SUBJECT'S LAST NAME:
VERIFICATION OF OIL
SUBJECT'S FIRST NAME: A
O.L. #: A
DEPT/AGENCY: DMED
CASE/REPORT: A
TEST TYPE: V
ALCO TARGET VALUE: .079
ALCO S/N: X301548

--- BREATH ANALYSIS ---

.079 ADJUSTED FOR 30.69 in
ALCO TARGET .081 13:33
BLANK TEST .000 13:34
INTERNAL STANDARD VERIFIED 13:34
ALCO TV 30.69 in .079 13:35
BLANK TEST .000 13:35
SUBJECT SAMPLE .079 13:36
BLANK TEST .000 13:37
ALCO TV 30.69 in .079 13:37
BLANK TEST .000 13:38

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130289

JANUARY 02, 2010
TIME 13:29

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 08/08/01
HEATERS
SAMPLE CHAMBER: 43c
BREATH TUBE: 41c
BAROMETER: 30.69 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@BCDEFG
HIJKLMNOPQRSTUVWXYZ[^_`abcde/ghijklmno
pqrstuvwxyz{|}~